Equivalent circuit models and analysis of electrochemical impedance spectra of caffeine solutions and beverages

Int. J. Electrochem. Sci., 11 **(2016)** 6370 – 6386

International Journal of ELECTROCHEMICAL SCIENCE
www.electrochemsci.org

Equivalent Circuit Models and Analysis of Electrochemical Impedance Spectra of Caffeine Solutions and Beverages

Gourav Bhattacharya[1], *Ashish Mathur*[2], *Srikanta Pal*[3] *James McLaughlin*[4] *and Susanta Sinha Roy*[1,*]

[1]Department of Physics, School of Natural Sciences, Shiv Nadar University, Gautam Budh Nagar 201314, Uttar Pradesh, India.
[2]Amity Institute of Nanotechnology, Amity University, Sector-125, Noida 201303, Uttar Pradesh, India.
[3]Department of Electrical Engineering, School of Engineering, Shiv Nadar University, Gautam Budh Nagar 201314, Uttar Pradesh, India.
[4]Nanotechnology and Integrated Bioengineering Centre, Jordanstown Campus, University of Ulster, Newtownabbey, BT37 OQB, Northern Ireland, United Kingdom.
[*]E-mail: susanta.roy@snu.edu.in

doi: 10.20964/2016.07.86

Received: 24 November 2015 / *Accepted:* 18 May 2016 / *Published:* 4 June 2016

In this study, a non-faradaic electrochemical impedance spectroscopy has been employed for estimation of caffeine concentration in beverages. Impedance spectra were recorded by using a two electrode system without adding any redox reagents in the measured solutions. Electrochemical impedance data of caffeine solutions in pure water and beverages were measured and an appropriate equivalent electrical circuit model is developed to help in this investigative analysis. The interaction of caffeine molecules with the electrodes was primarily correlated to the formation of electrical double-layer at modified interface. Overall system impedance (|Z|), inverse of solution resistance ($1/R_S$) and constant phase element of the system were further investigated from the equivalent electrical circuit and plotted as a function of pure caffeine concentration. Finally, the results obtained from spiked diluted cola beverages were compared with the pure caffeine sample.

Keywords: Caffeine, Impedance Spectroscopy, Circuit modeling

FULL TEXT